


<b>Name of the equipment</b>	<b>Climatic Test Chamber Make Tenny , Model No TJR</b>
<b>Purpose/Use</b>	<b>Used for conducting temperature test</b>
<b>Photograph</b>	
<b>Key specification</b>	<b>Temp range -60 to + 175 degree C</b>  <b>Rate of change of temperature 5 degree C/Min</b>
<b>Representative output, if possible</b>	<b>Cyclic wave</b>