

दूरभाष /Phone : 011-23905170

फैक्स /Fax : 011-23919509

ई-मेल /E-Mail : ppe @ inmas.drdo.in



पत्रांक: INMAS/TC/2518/COVID-19/2020

भारत सरकार, रक्षा मंत्रालय

GOVERNMENT OF INDIA, MINISTRY OF DEFENCE

रक्षा अनुसंधान एवं विकास संगठन

DEFENCE RESEARCH & DEVELOPMENT ORGANISATION

नाभिकीय औषधि तथा संबद्ध विज्ञान संस्थान

INSTITUTE OF NUCLEAR MEDICINE & ALLIED SCIENCES

ब्रिगेडियर एस. के. मजूमदार मार्ग, दिल्ली - 110054

BRIG. S.K. MAZUMDAR MARG, DELHI-110054

दिनांक /Date : 7<sup>th</sup> May, 2020

## UNIQUE CERTIFICATION CODE (UCC)

### for Personal Protection Equipment (PPE) Body Coverall for COVID-19

This Unique Certification Code is issued to **M/s Ajit Industries Pvt. Ltd**, with registered address as,

414/415, 4<sup>th</sup> floor, D Mall, Netaji  
Subhash Place, Pitampura,  
Delhi - 110034

Ref.: Affidavit No. IN-DL41717986145903S, Notary Attested  
Dated 06/05/2020

against their **Cover all nonwoven with blue tape**, test sample with Sample No. Nil, submitted to us on 13<sup>th</sup> April, 2020 and tested on 16<sup>th</sup> April, 2020. The sample has **passed** the **Synthetic Blood Penetration Test** conducted as per ISO 16603 class 3 and conforms to the technical requirement of Ministry of Health & Family Welfare guidelines dated 2<sup>nd</sup> March, 2020. The test results are annexed to this Certificate.

The Sample's UCC code is **COVID19/INMAS/160420-05-AJT-01**  
which is valid for **6 month(s)** from the date of issue.

The Affidavit/Declaration furnished by the applicant in relation to their sample is annexed to this Certificate.

This Certificate is based on the sample submitted by the applicant. No other test/verification has been conducted by INMAS other than results in the test report. The procurement agency is advised to conduct prior due diligence before placement of supply agreement, as per their procurement policy, including periodical sample tests during the course of supply of material.

This UCC has been issued with reference to Ministry of Textiles letter No F. No 8/4/2020-R&D dated 16-04-2020. The procedures enumerated at Para-3 and 4 of the instructions of Ministry of Textiles, Government of India issued vide letter No. F.No. 8/4/2020-R&D dated 6<sup>th</sup> April, 2020 must be strictly complied with regards to use of this certification.

-sd-

**Authorised Signatory**

Website: <https://www.drdo.gov.in/testing-and-certification-samples-ppe-body-coverall-covid-19-inmas>

Copy to HLL

*This is a soft copy of the Certificate and does not require signature. The original Certificate shall have a hologram and carry the authorized signatory's signature and official seal on all the pages*

दूरभाष /Phone : 011-23905170  
फैक्स /Fax : 011-23919509  
ई-मेल /E-Mail : ppe @ inmas.drdo.in



पत्रांक: INMAS/TC/2518/COVID-19/2020  
भारत सरकार, रक्षा मंत्रालय  
GOVERNMENT OF INDIA, MINISTRY OF DEFENCE  
रक्षा अनुसंधान एवं विकास संगठन  
DEFENCE RESEARCH & DEVELOPMENT ORGANISATION  
नाभिकीय औषधि तथा संबद्ध विज्ञान संस्थान  
INSTITUTE OF NUCLEAR MEDICINE & ALLIED SCIENCES  
ब्रिगेडियर एस. के. मजूमदार मार्ग , दिल्ली -110054  
BRIG. S.K. MAZUMDAR MARG, DELHI-110054

दिनांक /Date : 7<sup>th</sup> May, 2020

**M/s Ajit Industries Pvt. Ltd**

(Sample Tested at :R.H65%+/-2% and temp.21 Degree C+/-1Degree C+/-1 Degree C)

<b>Synthetic Blood Penetration Test ISO 16603 class III</b>	<b>Described by the Customer: Coverall Body Suit sample</b>
Result at Fabric Portion	Pass
Result at Seam Portion	Pass

**This UCC has been issued with reference to Ministry of Textile letter No F. No 8/4/2020-R dated 16-04-2020.**

Note COVID 19 warning-The above report pertains only to the sample submitted by the customer and the samples are not drawn by INMAS, DRDO. These results need not be indicative of the results of bulk lots of the articles under consideration.

End of Report