दूरभाष /Phone: 011-23905170 फैक्स /Fax: 011-23919509

ई-मेल /E-Mail: ppe @ inmas.drdo.in



पत्रांक: INMAS/TC/2518/COVID-19/2020 भारत सरकार, रक्षा मंत्रालय GOVERNMENT OF INDIA, MINISTRY OF DEFENCE रक्षा अनुसंधान एवं विकास संगठन DEFENCE RESEARCH & DEVELOPMENT ORGANISATION नाभिकीय औषधि तथा संबद्ध विज्ञान संस्थान INSTITUTE OF NUCLEAR MEDICINE & ALLIED SCIENCES ब्रिगेडियर एस. के. मज़मदार मार्ग, दिल्ली -110054

BRIG. S.K. MAZUMDAR MARG, DELHI-110054

Dated: 22<sup>nd</sup> June, 2020

## **UNIQUE CERTIFICATION CODE (UCC)**

for Personal Protection Equipment (PPE) Body Coverall for COVID-19

This Unique Certification Code is issued to M/s Clear Vision Outdoor Advertisers, with registered address as,

10059, Zameer Street, Nawab Ganj, Ref. Affidavit No. IN-DL44851032435157S, Notary Attested, Delhi-11006, DELHI-52 Dated 30/05/2020

against their **Coverall, white without tape, non-woven** test sample with Sample No. Nil 15/04/2020 submitted to us and tested on 30.04.2020. The sample has **passed** the **Synthetic Blood Penetration Test** conducted as per ISO 16603 class 3 and conforms to the technical requirement of Ministry of Health & Family Welfare guidelines dated 2<sup>nd</sup> March, 2020. The test results are annexed to this Certificate.

The Sample's UCC code is: COVID19/INMAS/300420-06/CVOA-01

which is valid for 6 months from the date of issue.

The Affidavit/Declaration furnished by the applicant in relation to their sample is annexed to this Certificate.

This Certificate is based on the sample submitted by the applicant. No other test/verification has been conducted by INMAS other than results in the test report. The procurement agency is advised to conduct prior due diligence before placement of supply agreement, as per their procurement policy, including periodical sample tests during the course of supply of material.

This UCC has been issued with reference to Ministry of Textiles letter No F. No 8/4/2020-R&D dated 16-04-2020. The procedures enumerated at Para-3 and 4 of the instructions of Ministry of Textiles, Government of India issued vide letter No. F.No. 8/4/2020-R&D dated 6<sup>th</sup> April, 2020 must be strictly complied with regards to use of this certification.

-sd-**Authorised Signatory** 

 $Website: {\it https://www.drdo.gov.in/testing-and-certification-samples-ppe-body-coverall-covid-19-inmas} \ Copy \ to \ HLL$ 

दूरभाष /Phone : 011-23905170 फैक्स /Fax : 011-23919509

ई-मेल /E-Mail: ppe @ inmas.drdo.in



पत्रांक: INMAS/TC/2518/COVID-19/2020 भारत सरकार, रक्षा मंत्रालय GOVERNMENT OF INDIA, MINISTRY OF DEFENCE रक्षा अनुसंधान एवं विकास संगठन DEFENCE RESEARCH & DEVELOPMENT ORGANISATION नाभिकीय औषधि तथा संबद्ध विज्ञान संस्थान INSTITUTE OF NUCLEAR MEDICINE & ALLIED SCIENCES ब्रिगेडियर एस. के. मजूमदार मार्ग, दिल्ली -110054 BRIG. S.K. MAZUMDAR MARG, DELHI-110054

दिनांक /Date: 22<sup>nd</sup> June, 2020

## M/s Clear Vision Outdoor Advertisers,

(Sample Tested at :R.H65%+/-2% and temp.21 Degree C+/-1Degree C+/-1 Degree C)

Synthetic Blood Penetration Test ISO 16603 class III	Described by the Customer: Coverall, white without tape, non-woven
Result at Fabric Portion	Pass
Result at Seam Portion	Pass

## This UCC has been issued with reference to MoT letter no F. No 8/4/2020-R dated 16-04-2020.

Note COVID 19 warning-The above report pertains only to the sample submitted by the customer and the samples are not drawn by INMAS, DRDO. These results need not be indicative of the results of bulk lots of the articles under consideration,

End of Report