दूरभाष /Phone: 011-23905170 फैक्स /Fax : 011-23919509

ई-मेल /E-Mail: ppe @ inmas.drdo.in



पत्रांक: INMAS/TC/2518/COVID-19/2020

भारत सरकार, रक्षा मंत्रालय

GOVERNMENT OF INDIA, MINISTRY OF DEFENCE

रक्षा अनुसंधान एवं विकास संगठन

DEFENCE RESEARCH & DEVELOPMENT ORGANISATION नाभिकीय औषधि तथा संबद्ध विज्ञान संस्थान INSTITUTE OF NUCLEAR MEDICINE & ALLIED SCIENCES

ब्रिगेडियर एस. के. मजूमदार मार्ग , दिल्ली -110054 BRIG. S.K. MAZUMDAR MARG, DELHI-110054

दिनांक /Date: 27<sup>th</sup> May, 2020

## **UNIQUE CERTIFICATION CODE (UCC)**

for Personal Protection Equipment (PPE) Body Coverall for COVID-19

This Unique Certification Code is issued to M/s Krishna Meditech, with registered address as,

5296, Old Anaj Mandi, Science Market, Ambala Cantt-133001, Haryana Ref. Affidavit No.AA282020D20, Notary Attested, Dated 28/04/2020

against their Coverall white with blue tape, Non-Laminated, Non- Woven, test sample with Sample No. Nil & Dated: 24<sup>th</sup> April 2020, submitted to us and tested on 25<sup>th</sup> May, 2020. The sample has passed the Synthetic Blood Penetration Test conducted as per ISO 16603 class 3 and conforms to the technical requirement of Ministry of Health & Family Welfare guidelines dated 2<sup>nd</sup> March, 2020. The test results are annexed to this Certificate.

The Sample's UCC code is: COVID19/ INMAS/250520-06/KMI-01

which is valid for **6 month(s)** from the date of issue

The Affidavit/Declaration furnished by the applicant in relation to their sample is annexed to this Certificate.

This Certificate is based on the sample submitted by the applicant. No other test/verification has been conducted by INMAS other than results in the test report. The procurement agency is advised to conduct prior due diligence before placement of supply agreement, as per their procurement policy, including periodical sample tests during the course of supply of material.

This UCC has been issued with reference to Ministry of Textiles letter No F. No 8/4/2020-R&D dated 16-04-2020. The procedures enumerated at Para-3 and 4 of the instructions of Ministry of Textiles, Government of India issued vide letter No. F.No. 8/4/2020-R&D dated 6<sup>th</sup> April, 2020 must be strictly complied with regards to use of this certification.

-sd-**Authorised Signatory** 

Website: https://www.drdo.gov.in/testing-and-certification-samples-ppe-body-coverall-covid-19-inmas

Copy to HLL

दूरभाष /Phone: 011-23905170 फैक्स /Fax : 011-23919509

ई-मेल /E-Mail: ppe @ inmas.drdo.in



पत्रांक: INMAS/TC/2518/COVID-19/2020

भारत सरकार, रक्षा मंत्रालय

GOVERNMENT OF INDIA, MINISTRY OF DEFENCE

रक्षा अनुसंधान एवं विकास संगठन

DEFENCE RESEARCH & DEVELOPMENT ORGANISATION नाभिकीय औषधि तथा संबद्ध विज्ञान संस्थान

INSTITUTE OF NUCLEAR MEDICINE & ALLIED SCIENCES ब्रिगेडियर एस. के. मजूमदार मार्ग , दिल्ली -110054

।श्रेगाडयर एस. क. मजूमदार माग , १६९९६ - । १००५ BRIG. S.K. MAZUMDAR MARG, DELHI-110054

दिनांक /Date: 27<sup>th</sup> May, 2020

M/s Krishna Meditech,

(Sample Tested at :R.H65%+/-2% and temp.21 Degree C+/-1Degree C+/-1 Degree C)

| Synthetic Blood Penetration Test ISO 16603 class III | Described by the Customer: Coverall Body Suit sample |
|--|--|
| Result at Fabric Portion                             | Pass   |
| Result at Seam Portion                               | Pass   |

## This UCC has been issued with reference to Ministry of Textile letter No F. No 8/4/2020-R dated 16-04-2020.

Note COVID 19 warning-The above report pertains only to the sample submitted by the customer and the samples are not drawn by INMAS, DRDO. These results need not be indicative of the results of bulk lots of the articles under consideration.

End of Report